



MOBILE APP GUIDE

Wearers



WHAT'S NEW IN THIS VERSION...

Instadose® Mobile App enhancements include new screens, dose summaries and analytics data, and improved functionality affording users a new level of actionable information, user engagement and compliance.

- ✓ **Device Scanning** from login screen (without the need for user login) with the ability to read dosimeters within range in the background.
- ✓ **Welcome Screen** with alert notifications front-and-center.
- ✓ **Multiple Accounts/Locations Access (for Admins Only)** allows admins to switch between multiple locations to maintain a global view of their entire dosimetry monitoring program and stay abreast of critical program stats.
- ✓ **Wearer Insights Screen** provides informative summary and trending data, and dose analytics based on user access level (Account/Location Administrator, Wearer).
- ✓ **Badges Screen (for Admins Only)** provides list of all assigned badges for selected Location enabling detailed drill-down view into wearer, badge, and dose data information.
- ✓ **My Badge Screen (for Wearers)** features detailed wearer information, dose overview, and read history results.
- ✓ **Badge Reader Screen** automatically searches for and transmits dose data from badges within range; provides visual confirmation of assigned badges that have communicated successfully (based on user login).
- ✓ **Activity Screen** provides actionable access to alerts (including high dose, communication overdue, and temperature error) with visibility into successful/unsuccessful dose reads.

ABOUT US...

Mirion Technologies Dosimetry Services Division (DSD) is the first name in dosimetry service innovation. We don't just deliver cutting edge products and services, we invent them. With over 45 years of dosimetry experience, we provide the most technologically advanced dosimeters on the market and a first-rate commitment to customer satisfaction. Our line of revolutionary Instadose® dosimeters are transforming the way facilities manage their radiation monitoring programs.

Instadose is a smarter radiation monitoring platform that is simplifying administration, reducing costs and transforming how staff are monitored and safeguarded from radiation over-exposure. Now immediate dose reads can be captured by internet-enabled computers, smart phones and tablets (Android and iOS), InstaLink-USBs, and InstaLink Hotspots.

The Instadose® Mobile App wirelessly communicates and securely transmits accumulated dose data from Instadose®+ and Instadose®2 digital dosimeters to Mirion servers where dose information is archived in the wearer's legal dose of record. Automatic dose reads are performed on an established calendar setting (set by your account administrator) and manual dose reads can be performed at any time by pressing the button on the back of the dosimeter for 4-seconds. Successful transmission occurs when the dosimeter is in range of the mobile device with the Instadose Mobile App opened and "searching for devices" appears on the main screen. Dose results, history, and wearer/dosimeter information can be immediately viewed (by the individual wearer or admin only) directly on the app after a successful dose communication/transmission.

User Access Levels

The Instadose Mobile App allows four user types to access applicable account and dose data:

- Radiation Safety Officer (RSO) is defined as the Account Administrator and can access info/details for all accounts, locations, and wearers.
- Location Administrators can access only the info/details for their assigned locations and wearers.
- Group Administrators can access only the info/details for assigned wearers within their group.
- Wearers can only access their own information and badge/dose details.**

Mobile App Accessibility

- An internet connection must be active on the mobile device.
- Bluetooth must also be enabled on the mobile device.
- For Android: Location Services must be enabled and the APP must be allowed to access the user location (in order for Bluetooth scanning to be enabled).
- Appropriate messages are displayed if any of the above accessibility conditions are not met.
- If all app accessibility conditions above are met, any badge within range can be read by the app on your mobile device.

Access Features for Wearers

Wearers can:

- ✓ Scan & read their assigned badges.
- ✓ View their own profile.
- ✓ Change their own username, password & email.
- ✓ View their wearer monthly dose.
- ✓ View their dose accumulation: this month/month-to-date, previous month, quarter-to-date, year-to-date and lifetime dose.
- ✓ View dose read history for all assigned badges.
- ✓ View when any assigned badge was last read.
- ✓ View their activity summary & badge alerts.


Instadose® Mobile App Wearer Guide

5

SIGN IN SCREEN

To Sign In:

- ❑ Enter Username & Password
- ❑ Click **'Sign In'**
- ❑ Checking **'Keep me signed in'** will enable the app to remember your sign in credentials.
- ❑ **'Sync devices automatically'** should be turned ON whenever possible, but specifically when scanning and reading devices is intended.



Instadose® Mobile APP Wearer Guide

6

WELCOME SCREEN

The welcome screen notifies the user of any badge alerts (high dose, communication overdue) registered within the last 30 days.

If there are any badge alerts, the alert icon color is orange.

Clicking on the alert icon will take the user to the Activity detail screen corresponding to the alert.



Instadose® Mobile APP Wea

7

OVERVIEW: Dose Card

Location **Dose Card** ring color

- Gray – If the current month is 0 or no dose history.
- Blue – If the current month has data/dose value.
- Orange – If the current month dose is higher than the previous month dose.

Dose Card displays the total accumulated dose value for current month (top) and last/previous month (bottom).

- '0' dose means - The dose exists but dose value is less than minimum reportable dose (MRD) value.
- 'No Dose History' means - NO dose record exists for that location for that month.

The current and last month records in dose card correspond to the values in the **Acc. Location Dose/Month** chart.

Results shown correspond to wearer profile/login entered (circled in red.)



Instadose® Mobile APP Wearer Guide

8

OVERVIEW: Welcome Tile - Alert Icon


There are two (2) types of alerts that, if applicable to the wearer/badge, will be displayed in the Welcome tile of the Overview screen:

- High Dose
- Communication Overdue

If the wearer has any badge alerts, the alert icon will be orange. Clicking on the alert icon will take the wearer to the detail screen corresponding to the alert.

For Communication Overdue Alerts:

- If the wearer has only one (1) assigned badge, the alert will prompt the wearer to read their badge - Clicking on this alert will take the wearer to the **Badge Reader** screen to read their assigned badge. (See page 23.)
- If the wearer has more than one (1) assigned badge, the alert will show the wearer the number of badges that need to be read - Clicking on this alert will take the wearer to the **Overdue Devices** detail screen listing all their assigned badges that have had unsuccessful dose reads (both automatic calendar set and manual) within the last 30-days.




Instadose® Mobile APP Wearer Guide

OVERVIEW: Welcome Tile - Alert Icon

For High Dose Alerts:

- If the wearer has only one (1) assigned badge, the alert will tell the wearer their badge has a recent high dose read - Clicking on this alert will take the wearer to the **Read History** detail screen showing all high dose alerts within last 30 days for their assigned badge.
- If the wearer has more than one (1) assigned badge, the alert will show the wearer the number of assigned badges that have recent high dose reads - Clicking on this alert will take the wearer to the **Read History** detail screen listing all their assigned badges with high dose reads within the last 30-days.



Instadose® Mobile APP Wearer Guide

OVERVIEW: My Average Dose/Month

My Average Dose/Month displays the wearer's total accumulated dose value by month.

Only months with existing data will be displayed on the graph, months with no existing data will be hidden.

The chart captures data for 12 last months.

If there is no data the 'No Dose History' message will be displayed.

The measurement displayed (mrem) corresponds to the Unit of Measure (UOM) set-up for the location assigned.



Instadose® Mobile APP Wearer Guide

11

OVERVIEW: My Average Dose Month (drill-down)

Clicking the **My Average Dose/ Month** allows the wearer to pick a Month & Year from the dropdown list (from 2000 to present yr).


Entering a Month & Year will take the wearer to the **Read History** screen listing their reads for the time period entered.

Note: The most recent reads, up to a maximum of 250 will be displayed for the selected period.

Results shown correspond to badges assigned to the wearer.

No data is displayed if there are no records to display (ex. new wearer with no prior data/records).

The measurement displayed (mrem) corresponds to the Unit of Measure (UOM) set-up for the location assigned.



Instadose® Mobile APP Wearer Guide


12

OVERVIEW: Last Badge Read

Last Badge Read shows the wearer's last badge read date, time and dose value.

The **Last Badge Read** tile also shows the wearer's cumulative dose values for:

- Month to Date
- Year to Date
- Quarter to Date
- Lifetime Dose



Instadose® Mobile APP Wearer Guide

13

MY BADGE SCREEN

The **My Badge** screen loads all the badges assigned to the wearer.

Full name of the wearer is displayed, followed by the body region for the badge (ex. Torso, Collar).



Instadose® Mobile APP Wearer Guide

14


MY BADGE SCREEN: Wearer Information & Dose Overview

Wearer Information displays:

- Photo of the badge type (name of the badge type appears under wearer name)
- Wearer full name
- Location (assigned)
- Role (corresponds to the role assigned/entered: *RadTech, Nurse, Cardiologist, Surgeon, etc.*)
- Group (corresponds to the assigned group)
- Body region (where badge is worn)
- Badge #
- Badge color

Dose Overview displays:

- Deep, Shallow and Eye dose for the Month, Quarter, Year and Lifetime to Date shown for the selected badge.
- The measurement displayed (mrem) corresponds to the Unit of Measure (UOM) set-up for the location/account assigned.




Instadose® Mobile APP Wearer Guide

MY BADGE SCREEN: Wearer Read History

Clicking the **Read History** tile will take wearer to their **Read History** screen.

Read data displayed includes:

- Dose read date/time
 - Type of read (calendar or manual)
 - Body Region (where the device is worn)
 - Measurement displayed (mrem) corresponds to the Unit of Measure (UOM) set-up for the location assigned.
- Note: An *(asterisk) will be displayed if the dose value is less than Minimum Reportable Dose (MRD).*
- Temp Error (unsuccessful dose reads)




Instadose® Mobile APP Wearer Guide

MY BADGE SCREEN: Wearer Read History (Drill-down)

Clicking on a dose record from the **Read History** screen will take the user to the **Read History** detail screen for the selected dose record.

Read data displayed includes:

- Dose read date/time
- Type of read (calendar or manual)
- Body Region (where the device is worn)
- Deep, Shallow, Eye dose values.
*Note: An *(asterisk) will be displayed if the dose value is less than Minimum Reportable Dose (MRD).*
- Measurement displayed (mrem) corresponds to the Unit of Measure (UOM) set-up for the location assigned.



BADGE READER STATUS

The octagon-shaped status icon turns blue when badges have been detected.

Status is shown (in blue text) below the octagon icon.

There are 8 status types:

- Looking for badges
- Connecting badge
- Reading
- Querying (serial#)
- Reading doses (serial#)
- Sending reads for badge (serial#)
- Finished reading badge (serial#)
Note: this is a successful read.
- Time Out/Unsuccessful
Note: this is an unsuccessful read.



BADGE READER SCREEN

After **successful communication**, the details shown correspond to the badge read.


- Badge type (ex. Instadose+, Instadose2)
- Serial number (in blue)
- Type of read (calendar/manual)
- Date/time of read
- Dose value (deep)

An *(asterisk) will be displayed if the dose value is less than Minimum Reportable Dose (MRD).

Measurement displayed (mrem) corresponds to the Unit of Measure (UOM) set-up for the location assigned.

If multiple dose records (stored on the dosimeter) are communicated at the same time, only the last date/time stamp will show on this screen. Click Read History to view all dose records communicated/transmitted.

Clicking on the dose record will take wearer to the **Read History** detail screen for that dose record.




Instadose® Mobile App Wearer Guide

BADGE READER SCREEN

If there is a **Temp Error**, the following details will be displayed:

- Badge type (ex. Instadose+, Instadose2)
- Serial number (in blue)
- Type of read (calendar/manual)
- Date/time of read
- Temp Error (where the dose value should be)

Clicking on the temperature error dose record will reveal the **'Temperature Error!'** pop-up message instructing users what to do next.




Instadose® Mobile App Wearer Guide

ACTIVITY SCREEN

The **Activity** screen displays all applicable alerts for badges assigned to the wearer:

- High Dose
- Communication Overdue
- Temperature Error



Instadose® Mobile APP Wearer Guide


21

ACTIVITY SCREEN: High Dose Alerts

Clicking the **High Dose** alert will take the wearer to the **Read History** screen showing a list of recent high dose reads (within last 30 days) along with corresponding details:

- Date and time of read
- Wearer name – body region
- Dose details (deep/shallow/eye)
- Measurement displayed (mrem) corresponds to the Unit of Measure (UOM) set-up for the location assigned.

Clicking on any of the dose records listed (on the **Read History** screen) will take the wearer to the **My Badge** screen showing dose details for the selected dose record.




Instadose® Mobile APP Wearer Guide

22

ACTIVITY SCREEN: Communication Overdue (single badge)

If a wearer has only one (1) badge assigned to them, clicking the **Communication Overdue** tile will take the wearer to the **Badge Reader** screen, with a pop-up message instructing the wearer how to read their badge.

Communication Overdue alerts can occur when either an automatic calendar read or manual read is not successfully performed/transmitted.



Instadose® Mobile APP Wearer Guide

23


ACTIVITY SCREEN: Communication Overdue (multiple badges)

If wearer has multiple badges assigned to them, clicking the **Communication Overdue** tile will take the wearer to an **Overdue Devices** detail screen listing any badges assigned to the wearer, which are overdue for communication.

- Date and time of read
- Wearer Name – Body Region
- Badge serial # (in blue)

Clicking on any of the badges listed (on the **Overdue Devices** screen) will take the wearer to the **Badge Reader** screen, with a pop-up message instructing the wearer how to read their badge.

Communication Overdue alerts can occur when either an automatic calendar read or manual read is not successfully performed/transmitted.




Instadose® Mobile APP Wearer Guide

24

ACTIVITY SCREEN: Temperature Error

Clicking the **Temp Error** tile will take wearer to **Read History** detail screen showing all their devices with recent temperature error reads.

- Date and time of read
- Wearer Name – Body Region
- Temp Error (in orange)




Instadose® Mobile APP Wearer Guide

25

ACTIVITY SCREEN: Temperature Error Message

Clicking on a wearer device with a temperature error will reveal the **'Temperature Error!'** pop-up message instructing users what to do next.



Instadose® Mobile APP Wearer Guide


26

ABOUT SCREEN

The **About** screen provides information about the product, company and Mobile App.

The new Instadose Mobile App (version 2.0.0) is compatible with:

- iOS version iOS 9.0 to 12
- Android version Kitkat 4.4 to Pie 9



Instadose® Mobile App Wearer Guide

USER PROFILE SCREEN


Clicking the triangle icon next to the wearer name will show profile details corresponding to the wearer/user log-in.

The **User Profile** screen provides user details, including:

- Username
- Email Address
- Account
- Account Role
- Locations (wearer location scope)

The following menu options are also available from this screen:


- About (information about the app)
- Change Password
- Sign Out



Instadose® Mobile APP Wearer Guide

USER PROFILE: Change Username

Allows username to be changed.




Instadose® Mobile APP Wearer Guide

29

USER PROFILE: Change Email Address

Allows email address to be changed.



Instadose® Mobile APP Wearer Guide

30

USER PROFILE: Change Password

Allows user to change password.

Password must contain:

- Minimum of 8 characters
- One uppercase letter (A-Z)
- One lowercase letter (a-z)
- One number (0-9)
- One special character (@+!/!%&*_#^.;:.,<>?=-~-)

Instadose® Mobile APP Wearer Guide

31

Warning Message: Bluetooth is OFF


When Bluetooth is OFF, the orange message text (pictured) will appear at the bottom of the screen.

Instadose® Mobile APP Wearer Guide

32

Warning Message: Internet is OFF

When there is NO Internet connection, the orange message text (pictured) will appear at the bottom of the screen.




Instadose® Mobile APP Wearer Guide

33

Warning Message: Internet & Bluetooth OFF

When both the Internet Connection and Bluetooth are OFF, the orange message text (pictured) will appear at the bottom of the screen.




Instadose® Mobile APP Wearer Guide

34

Info/Warning Message: Bluetooth turned ON without an active internet connection.

When Bluetooth is turned ON, but there is no internet connection, the Instadose Mobile App will return the message (pictured in the notification window on the device's home screen) to notify user that badge scanning is not possible without an internet connection.




Instadose® Mobile APP Wearer Guide

35

Info/Warning Message: When Instadose is running in the background.

When the Instadose Mobile App is running in the background, the message (pictured in the notification window on the device's home screen) will appear to notify user that badge scanning/read transmission was successfully performed.



Instadose® Mobile APP Wearer Guide

36

Contact Us!



dsd-support@mirion.com



U.S./Canada +.800.251.3331

U.K. 0170.629.9329

Worldwide +1.949.419.1000



www.instadose.com

Copyright ©2019 Mirion Technologies, Inc. or its affiliates. All rights reserved. Mirion, the Mirion logo, and other trade names of Mirion products listed herein are registered trademarks or trademarks of Mirion Technologies, Inc. or its affiliates in the United States and other countries. Third party trademarks mentioned are the property of their respective owners.